



DTR Medical[®]

AN
INNOVIA
MEDICAL
COMPANY



Cervical Rotating Biopsy Punch Product Specification



The Design Checklist

- Latest generation top jaw now stronger than Titanium, and 4x the strength of previous jaw
- Allows a Biopsy to be taken from a Hardened Cervix
- Low profile jaw provides better access to the transformation zone when migrating into the Endocervical Canal
- 360° rotation enables enhanced positioning and patient interaction
- Ensures a sharp, clean cut reducing Cervical trauma and patient discomfort
- Ergonomic handle designed for comfort
- Provides value-based benefits, improving procedure quality, patient safety, and overall cost savings
- Lightweight compared to alternatives
- Ideal for busy Colposcopy Clinics, with high-quality samples taken first-time
- Useful for increased HPV testing demands
- Single-use eliminates cross-contamination, inadequate samples, and repeat procedures
- UK manufactured
- Sterile, individually packed

DTR Medical Ltd, 17 Clarion Court, Enterprise Park, Swansea SA6 8RF, UK
t: +44(0) 1792 797910 e: dtrmarketing@innoviamedical.com w: www.dtrmedical.com



YOUR SINGLE-USE SOLUTION WITH CONFIDENCE



Cervical Rotating Biopsy Punch Product Specification

Intended Use

The Cervical Rotating Biopsy Punch has been designed specifically for use in Gynaecological procedures to obtain a Cervical Biopsy.

Equipped with a pair of cupped jaws at the end of a metal rod with a handle and trigger. The long metal construction ensures better access to the Cervix for the jaws, whilst the handle is operated outside. When the trigger is pulled the sharp jaw closes, this action removes a segment of Tissue for the Biopsy.

It is a Single-use device which is disposed of once used.

Instructions For Use (IFU)

There is no IFU available for the Cervical Rotating Biopsy Punch.

Conformity to European Regulations

Conforms to Medical Directive 93/42/EEC.

Manufactured under EN ISO:13485 in an ISO 7 Cleanroom Environment.

Product Specification



Description	Pack Size	Box Size	Code
Cervical Rotating Biopsy Punch	1 Biopsy Punch	10 Units	CBR1041



Material Breakdown

Product Component	Details
Jaw	17-4 Precipitation Hardened Stainless Steel
Tube	316 Stainless Steel
Handle	Polycarbonate (PC)
Trigger	Polycarbonate (PC)
Rotation Controller	Polycarbonate (PC)